

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
FLUCONAZOLE FLU 0.016-256 µg/mL	062321156	23.06.2021	2024.06.22

Ref.

92147 – 921470 – 921471

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	128-≥256	≥256
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	1-8	4
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.125-0.5	0.5

Batch Release

Approved

Date

26.07.2021

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.