

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
FLUCONAZOLE FLU 0.016-256 µg/mL	010722014	07.01.2022	2025.01.06

Ref.

92147 – 921470 – 921471

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	128–≥256	≥256
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	1–8	2
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.125–0.5	0.5

Batch Release

Approved

Date

24.01.2022

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.