

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Itraconazole ITC 0.002-32 µg/mL	032720015	27.03.2020	2023.03.27

Ref.

92148 – 921480 – 921481

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation*	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24 h	0.25-1	0.38
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24 h	0.064-0.25	0.125
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24 h	0.064-0.25	0.125

*If growth is insufficient at 24 h, plates can be read after a total of 48 h of incubation

Batch Release

Approved

Date

09.04.2020

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.