

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
<b>Voriconazole VO</b> <b>0.002-32 µg/mL</b>	121818006	18.12.2018	2021.12.17

**Ref.**

92150 – 921500 – 921501

**Antimicrobial Susceptibility Testing**

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.25-1	0.25
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.016-0.064	0.023
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.004-0.016	0.006

**Batch Release**

Approved

**Date**

09.01.2019

**Signature**Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.