

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Amphotericin B AMB 0.002-32 µg/mL	100516037	05.10.2016	2019.10.05

Ref.

92153 – 921530 – 921531

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.25-1	0.38
<i>Candida krusei</i> ATCC® 6528	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.5-2	1.5
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.125-0.5	0.38

Batch Release

Approved

Date

20.10.2016

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.