

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Production date | Expiration date |
|-----------------------------------|-----------|-----------------|-----------------|
| Amphotericin B AMB 0.002-32 µg/mL | 082818077 | 28.08.2018 | 2021.08.27 |

Ref.

92153 – 921530 – 921531

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

| Control strains | Medium | Inoculum | Incubation | Expected Results MIC range (mg/L) | Results MIC(mg/L) |
|--|---------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Candida parapsilosis</i> ATCC® 22019 | R.P.M.I. Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 24-48 h | 0.25–1 | 0.75 |
| <i>Candida krusei</i> ATCC® 6258 | R.P.M.I. Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 24-48 h | 0.5–2 | 1.5 |
| <i>Candida albicans</i> ATCC® 90028 | R.P.M.I. Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 24-48 h | 0.125–0.5 | 0.5 |

Batch Release

Approved

Date

05.09.2018

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.