

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Caspofungin CAS 0.002-32 µg/mL	111617086	16.11.2017	2020.11.15

Ref.

92154 – 921540 – 921541

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents


Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.25-1	0.5
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.25-2	0.5
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.064-0.25	0.19

Batch Release

Approved

Date

20.12.2017

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.