

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Anidulafungin AND 0.002-32 µg/mL	071520034	15.07.2020	2023.07.15

Ref.

92155 – 921550 – 921551

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.5–4	2
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.016–0.125	0.032
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.002–0.008	0.006

Batch Release

Approved

Date

03.08.2020

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.