

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Vancomycin/Teicoplanin VA/TEC 0.5-32/0.5-32 µg/mL	120919124	09.12.2019	2022.12.08

Ref.

92163 – 921630 – 921631

Antimicrobial Susceptibility Testing

Mueller Hinton II Agar + 5% Sheep Blood, 0.5 McFarland in broth, ambient, 35 ± 2°C, 18-48 h

Control strains	Expected Results	Expected Results	Results
	Vancomycin MIC range (mg/L)	Teicoplanin MIC range (mg/L)	
<i>Staphylococcus aureus</i> ATCC® 29213 (negative control)	0.5–2	1–4	Conforms
<i>Staphylococcus aureus</i> ATCC® 700698 (h-GISA)	1–8	≥32	Conforms
<i>Staphylococcus aureus</i> ATCC® 700699 (GISA)	4–16	≥32	Conforms

Batch Release

Approved

Date

02.01.2020

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.