

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Meropenem/Meropenem+EDTA MRP/MRD 0.125-8/0.032-2 µg/mL	010317095	03.01.2017	2020.01.03

Ref.

92165 – 921650 – 921651

Antimicrobial Susceptibility Testing

Mueller Hinton II Agar, 0.5 McFarland in saline, ambient, 35 ± 2°C, 16-20 h
Synergy of meropenem with EDTA

Control strains	Expected Results Ratio of MRP/MRD*	Results
<i>Klebsiella pneumoniae</i> ATCC® BAA-2146 (MBL positive)	≥ 8	Conforms
<i>Pseudomonas aeruginosa</i> ATCC® 27853 (MBL negative)	< 8	Conforms

*Presence of a phantom zone or ellipse deformation indicate MBL production as well

Batch Release

Approved

Date

10.01.2017

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.