

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Amoxicillin*-clavulanic acid (2 µg/mL) AMC 0.016-256* µg/mL	040918020	09.04.2018	2021.04.08

Ref.

92180 – 921800 – 921801

Antimicrobial Susceptibility Testing

Tested according to current EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 1°C, ambient 16-20 h	2–8	4
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 1°C, ambient 16-20 h	4–32	4
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland in saline	35 ± 1°C, 5% CO ₂ 16-20 h	0.125–0.5	0.38

Batch Release

Approved

Date

02.05.2018

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.