

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Amoxicillin-clavulanic acid AMC 0.016/2-256/2 µg/mL	102521102	25.10.2021	2024.10.24

Ref.

92180 – 921800 – 921801

Antimicrobial Susceptibility Testing

Tested according to current EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 1°C, ambient 16-20 h	2–8	4
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 1°C, ambient 16-20 h	4–32	6
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland in saline	35 ± 1°C, 5% CO ₂ 16-20 h	0.125–0.5	0.38
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton Fastidious Agar	0.5 McFarland in saline	35 ± 1°C, 5% CO ₂ 16-20 h	0.03–0.12	0.03

Batch Release

Approved

Date

02.11.2021

SignatureQuality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.