

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Ampicillin*-sulbactam (4 µg/mL) SAM 0.016-256* µg/mL	110817031	08.11.2017	2020.11.07

Ref.

92181 – 921810 – 921811

Antimicrobial Susceptibility Testing

Tested according to current EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	2
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	16–128	24
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland in saline	35 ± 1°C, 5% CO ₂ 16-20 h	0.06–0.25	0.19

*Mueller Hinton II (Cation Adjusted) Agar: Liofilchem, ref. 10031; BD-BBL, ref. 212257.

Batch Release

Approved

Date

22.11.2017

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.