

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
<b>Ampicillin*-sulbactam (4 µg/mL) SAM 0.016-256* µg/mL</b>	061920009	19.06.2020	2023.06.19

**Ref.**

92181 – 921810 – 921811

**Antimicrobial Susceptibility Testing**

Tested according to current EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	3
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	16–128	32
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland in saline	35 ± 1°C, 5% CO <sub>2</sub> 16-20 h	0.06–0.25	0.125

\*Mueller Hinton II Agar (Cation Adjusted): Liofilchem ref. 10031; BD-BBL ref. 254032.

**Batch Release**

Approved

**Date**

13.07.2020

**Signature**Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.