

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:	Xpert® Carba-R	
Cepheid Catalog	gue Part No.: GXCAR	BAR-CE-10
Kit Lot No.:	1000067050	
Cartridge Lot No	o.: 06609	
Kit Expiration Γ	Date: 2018-06-24	

Legal Manufacturer

Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA

## Functional Testing

Test Description	Acceptance Criteria	Test Result
LOW Positive	IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
HIGH Positive	IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
Negative	IMP1 NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

		6	29/1	13/17
Signature	of Quality Assurance		Date	
Name:	Loan Hoang	,		
Title:	Quality Systems Specialist	•		