

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:_ Xpert® C. difficile BT		
Cepheid Catalogue Part No.: GXCDIFFBT-CE-10		
Kit Lot No.: 1000202651	17	
Cartridge Lot No.: 32604		
Kit Expiration Date: 2021-11-28		
Legal Manufacturer	Manufacturing Facility	○ Sunnyvale
Cepheid AB	Cepheid AB	
Röntgenvägen 5 SE-171 54 Solna	Röntgenvägen 5 SE-171 54 Solna	
Sweden	Sweden	
Functional Testing		
Test Description Accept	Acceptance Criteria	

Toxigenic C.diff POS;Binary Toxin POS;027 PRESUMPTIVE POS

Toxigenic C.diff NEG;Binary Toxin NEG;027 NEG

JAL ()	2020-06-09
Signature of Quality Assurance	Date

Name: Eda Kozanli

Positive Control

Negative Control

Title: QA Manager Release

Passed

Passed