

Signature of Quality Assurance

Name: Siri Kjerrgren

Title: QA Analyst

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:_ Xpert® C. difficile BT				
Cepheid Catalogue Part No.: GXCDIFFBT-CE-10				
Kit Lot No.: 10002	248545			
Cartridge Lot No.: 331	05			
Kit Expiration Date: 2022-08-21				
Legal Manufacturer		Manufacturing Facility	Solna	○ Sunnyvale
Cepheid AB		Cepheid AB	•	Ç,
Röntgenvägen 5		Röntgenvägen 5		
SE-171 54 Solna		SE-171 54 Solna		
Sweden		Sweden		
Functional Testing				
Test Description	Acceptance Criteria		Test Result	
Positive Control	Toxigenic C.diff POS;Binary Toxin POS;027 PRESUMPTIVE POS		Passed	
Negative Control	Toxigenic C.diff NEG;Binary Toxin NEG;027 NEG			Passed
)				,