

Negative Control

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:_	Xpert® C. difficile BT			
Cepheid Catalogue Pa	art No.: GXCDIFFBT-CE-10			
Kit Lot No.:	0263986			
Cartridge Lot No.: 33208				
Kit Expiration Date: 2022-11-06				
Legal Manufacturer		Manufacturing Facility	Solna	C Sunnyvale
Cepheid AB Röntgenvägen 5		Cepheid AB Röntgenvägen 5		
SE-171 54 Solna		SE-171 54 Solna		
Sweden		Sweden		
Functional Testing				Γ
Test Description	Acceptance Criteria			Test Result
Positive Control	Toxigenic C.diff POS;Binary Toxin POS;027 PRESUMPTIVE POS			Passed

Toxigenic C.diff NEG;Binary Toxin NEG;027 NEG

Signature of Quality Assurance

Name: Siri Kjerrgren

Title: QA Analyst

Passed