

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® HCV Viral Load				
Cepheid Catalogue Part No.: GXHCV-VL-CE-10				
Kit Lot No.: 1000181821				
Cartridge Lot No.: 33704				
Kit Expiration Date: 2021-05-09				
Legal Manufacturer		Manufacturing Facility	Solna	Sunnyvale
Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden		Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden	( Solid (	Jumyvaic
Sweden		Sweden		
Functional Testing according to D14282 Rev.: K				
Test Description	Acceptance Criteria			Test Result
LOW Positive	HCV DETECTED: 1.50-1.90 log <sub>10</sub> IU/mL			Passed
HIGH Positive	HCV DETECTED: 4.25-4.55 log <sub>10</sub> IU/mL			Passed
Negative	HCV NOT DETECTED			Passed
Signature of Quality Assurance Date				

A Manager Release