

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Facility Solna C Sunnyvale

Functional T	Testino	according	to D14	1282	Rev .	K

Test Description	Acceptance Criteria	Test Result
		i csi nesan
LOW Positive	HCV DETECTED: 1.50-1.90 log ₁₀ IU/mL	Passed
HIGH Positive	HCV DETECTED: 4.25-4.55 log10 IU/mL	Passed
Negative	HCV NOT DETECTED	Passed

Meulii 20200226 Date

Signature of Quality Assurance

Name: Kambiz Ghaderi

Title: QA Analyst

P/N 301-5503, Rev. D