

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Directive and the Canadian Medical Device Regulations (CMDR).			
Product Name: Xpert® HCV Viral Load			
Cepheid Catalogue Part No.: GXHCV-VL-CE-10			
Kit Lot No.: 1000186215			
Cartridge Lot No.: 33803			
Kit Expiration Date: 2021-06-06			
Legal Manufacturer	Manufacturing Facility	Sunnyvale	
Cepheid AB Röntgenvägen 5 SE-171 54 Solna	Cepheid AB Röntgenvägen 5 SE-171 54 Solna	,	
Sweden	Sweden		
Functional Testing acco	ording to D14282 Rev.: K		
Test Description	Acceptance Criteria	Test Result	
LOW Positive	HCV DETECTED: 1.50-1.90 log ₁₀ IU/mL	Passed	
HIGH Positive	HCV DETECTED: 4.25-4.55 log ₁₀ IU/mL	Passed	
Negative	HCV NOT DETECTED	Passed	
1/			
Kamber	Letti 20200220		
Signature of Quality Assurance Date			
Name: Kambiz Ghaderi			
Title: QA Analyst			