



CERTIFICATE OF ANALYSIS

KIT NAME: Cleaner for Mythic 22
CAT. NO: HM22-001-1
LOT NO: 909-30
EXPIRY DATE: 2021-09
DATE OF QC: 26.09.2019

<i>PARAMETER</i>	<i>NORM</i>	<i>RESULTS</i>
Appearance	clear	clear
Colour	violet	violet
pH (25 ± 0.5°C)	8.40 – 8.60	8.40
Conductivity [mS/cm]	14.20 – 15.20	14.40
Osmolality [mOsmol/kg H ₂ O]	1880 - 1920	1898
Background:		
WBC K/μl	<0.3	<0.3
RBC M/μl	<0.05	<0.05
HGB g/dl	<0.3	<0.3
PLT K/μl	<5	<5
Functional test	correct	correct

We hereby certify that this lot has been tested and was found to meet our specifications.

Sign.: Anna Spryszak

Quality Control Department

This certificate is a computer print out and has therefore not been signed by hand.