



CERTIFICATE OF ANALYSIS

KIT NAME:	Diluent for Mythic 22
CAT. NO:	HM22-003-10
LOT NO:	909-97
EXPIRY DATE:	2021-09
DATE OF QC:	02.10.2019

<i>PARAMETER</i>	<i>NORM</i>	<i>RESULTS</i>
Appearance	clear	clear
Colour	colourless	colourless
Odour	characteristic odour of 2-phenoxyethanol	characteristic odour of 2-phenoxyethanol
pH (25 ± 0.5°C)	7.35 – 7.55	7.38
Conductivity [mS/cm]	17.00 – 18.00	17.89
Osmolality [mOsmol/kg H ₂ O]	331 - 351	344
Background:		
WBC K/μl	<0.3	<0.3
RBC M/μl	<0.05	<0.05
HGB g/dl	<0.3	<0.3
PLT K/μl	<5	<5
Functional test	correct	correct

We hereby certify that this lot has been tested and was found to meet our specifications.

Sign.: Anna Spryszak
Quality Control Department

This certificate is a computer print out and has therefore not been signed by hand.

19, chemin du Champ-des-Filles – CH-1228 PLAN-LES-OUATES / GENEVA / SWITZERLAND
Phone : +4122 884 90 90 – Fax : +4122 884 90 99
WEB : www.orphee-medical.com - Email : contact@orphee-medical.com