

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Ampicillin AMP 10 µg	090919082	2022.09.08

Ref.

9006 – 9006/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	15–22	18
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	27–35	29
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	6	6
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Medium	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-18 h	13–21	18
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar with 5% Sheep Blood	0.5 McFarland	35 ± 1°C, 5% CO ₂ 20-24 h	30–36	33

Batch Release

Approved

Date

12.09.2019

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.