

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Aztreonam ATM 30 µg	032621069	2024.03.25

Ref.

9008 – 9008/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	28-36	32
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	23-29	27
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	31-38	31
<i>Klebsiella pneumoniae</i> ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	10-16	12
<i>Klebsiella pneumoniae</i> ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	9-17*	12
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Medium	0.5 McFarland	35 ± 1°C, 5% CO2 16-18 h	30-38	32

*Established and validated by EUCAST

Batch Release

Approved

Date

07.04.2021

Signature

Quality Control
(D. Vitagliano)

The results reported were obtained at the time of release.