

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Expiration date |
|--------------------|-----------|-----------------|
| Cefazolin KZ 30 µg | 060818021 | 2021.06.07 |

Ref.

9015 – 9015/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains | Medium | Inoculum | Incubation | Expected Results Zone range (mm) | Results Zone (mm) |
|---|---------------------------|---------------|------------------------------|-------------------------------------|----------------------|
| <i>Escherichia coli</i> ATCC® 25922 | Mueller Hinton II Agar | 0.5 McFarland | 35 ± 1°C, ambient 16-18 h | 21–27 | 22 |
| <i>Staphylococcus aureus</i> ATCC® 25923 | Mueller Hinton II Agar | 0.5 McFarland | 35 ± 1°C, ambient 16-18 h | 29–35 | 31 |

Batch Release

Approved

Date

14.06.2018

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.