

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Cefuroxime CXM 30 µg	011020107	2023.01.09

Ref.

9021 – 9021/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	20-26	25
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	27-35	34
<i>Haemophilus influenzae</i> ATCC® 49766	Haemophilus Test Medium	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-18 h	28-36	29
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	26-34*	28
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	28-34*	30
<i>Neisseria gonorrhoeae</i> ATCC® 49226	GC Agar Base with 1% Defined Growth Suppl.	0.5 McFarland	36 ± 1°C, 5% CO ₂ 20-24 h	33-41	35

*Established and validated by EUCAST

Batch Release

Approved

Date

15.01.2020

Signature

Quality Control
(D. Vitagliano)

The results reported were obtained at the time of release.

Dario Vitagliano