

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Chloramphenicol C 30 µg	061319086	2022.06.12

Ref.

9022 – 9022/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	21-27	27
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	19-26	26
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	20-28*	26
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Medium	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-18 h	31-40	31
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar with 5% Sheep Blood	0.5 McFarland	35 ± 1°C, 5% CO ₂ 20-24 h	23-27	23
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	24-30*	24
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	31-37*	32


*Established and validated by EUCAST

Batch Release

Approved

Date

19.06.2019

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.