

# Liofilchem®

## Certificate of Analysis

Product	Batch	Expiration date
Gentamicin CN 10 µg	092818086	2021.09.27

**Ref.**

9026 – 9026/1

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	19-26	21
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	19–27	26
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	19–25*	25
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	16–21	21

\*Established and validated by EUCAST

**Batch Release**

Approved

**Date**

03.10.2018

**Signature**Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.