

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Neomycin N 30 µg	070620157	2023.07.06

Ref.

9032 – 9032/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	17-23	23
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	18-26	23

Batch Release

Approved

Date

14.07.2020

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.