

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Fusidic acid FC 10 µg	121719109	2022.12.16

Ref.

9049 – 9049/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	24–32	31
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	26–32*	29
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar with 5% Sheep Blood	0.5 McFarland	35 ± 1°C, 5% CO ₂ 20-24 h	9–16	16

*Established and validated by EUCAST

Batch Release

Approved

Date

03.01.2020

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.