

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Ciprofloxacin CIP 5 µg	101018076	2021.10.09

Ref.

9056 – 9056/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	30-40	34
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	29-37*	32
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	22-30	26
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	21-27*	25
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	25-33	33
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	19-25*	20
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Medium	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-18 h	34-42	35
<i>Neisseria gonorrhoeae</i> ATCC® 49226	GC Agar Base with 1% Defined Growth Suppl.	0.5 McFarland	36 ± 1°C, 5% CO ₂ 20-24 h	48-58	51
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	22-28*	23
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h	32-40*	33
<i>Campylobacter jejuni</i> ATCC® 33560	Mueller Hinton Fastidious Agar	0.5 McFarland	41 ± 1°C microaerobic 24 h	34-42*	36

*Established and validated by EUCAST

Batch Release

Approved

Date

15.10.2018

Signature

Quality Control
(D. Vitagliano)

The results reported were obtained at the time of release.

Dario Vitagliano