

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Amphotericin B AMB 20 µg	071618117	2021.07.15

Ref.

9071 – 9071/1

Antimicrobial Susceptibility Testing

Tested according to in-house disc diffusion method for antifungal agents

Control strains	Medium	Inoculum	Incubation	Results Halo (mm)
<i>Candida albicans</i> ATCC® 90028	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	16
<i>Candida parapsilosis</i> ATCC® 22019	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	11
<i>Candida krusei</i> ATCC® 6258	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	16
<i>Candida tropicalis</i> ATCC® 750	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	15

Batch Release

Approved

Date

26.07.2018

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.