

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Expiration date |
|--------------------------|-----------|-----------------|
| Amphotericin B AMB 20 µg | 010819030 | 2022.01.07 |

Ref.

9071 – 9071/1

Antimicrobial Susceptibility Testing

Tested according to in-house disc diffusion method for antifungal agents

| Control strains | Medium | Inoculum | Incubation | Results Halo (mm) |
|--|--|---------------|---------------------------------|----------------------|
| <i>Candida albicans</i> ATCC® 90028 | Mueller Hinton+ Glucose+ Methylene Blue Agar | McFarland 0.5 | Aerobic , 24-48 h , 35 ± 2°C | 17 |
| <i>Candida parapsilosis</i> ATCC® 22019 | Mueller Hinton+ Glucose+ Methylene Blue Agar | McFarland 0.5 | Aerobic , 24-48 h , 35 ± 2°C | 20 |
| <i>Candida krusei</i> ATCC® 6258 | Mueller Hinton+ Glucose+ Methylene Blue Agar | McFarland 0.5 | Aerobic , 24-48 h , 35 ± 2°C | 14 |
| <i>Candida tropicalis</i> ATCC® 750 | Mueller Hinton+ Glucose+ Methylene Blue Agar | McFarland 0.5 | Aerobic , 24-48 h , 35 ± 2°C | 18 |

Batch Release

Approved

Date

15.01.2019

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.