

Liofilchem S.r.l.	CERTIFICATO CONTROLLO QUALITÀ	N°9078 - 9078/1
	<i>QUALITY CONTROL CERTIFICATE</i>	Revisione 9 del 09.10.2013 Pag. 1 di 1

CARATTERISTICHE DEL PRODOTTO / PRODUCT CHARACTERISTICS

Prodotto / <i>Product</i>	NYSTATIN
Sigla / <i>Initial</i>	NY 100 IU
Lotto / <i>Batch</i>	021720096
Data produzione / <i>Production date</i>	17.02.2020
Data scadenza / <i>Expiry date</i>	2023.02.16
Contenuto antibiotico del disco / <i>Antimicrobial Disk content</i>	Nystatin 100 IU

CONTROLLO QUALITÀ / QUALITY CONTROL

Terreno di coltura / <i>Culture medium</i>	Mueller Hinton+Glucose+Methylen Blue Agar
Tempo di incubazione / <i>Incubation time</i>	20-24 h in caso di crescita insufficiente prolungare l'incubazione fino a 48 ore / <i>read at 48 hours only when insufficient growth is observed after 24 hours of incubation</i>
Temperatura / <i>Temperature</i>	36 ± 1°C
Modalità di incubazione / <i>Procedure of incubation</i>	O₂
Microrganismo/ <i>Microorganism</i>	RESULTS mm
<i>C. albicans</i> ATCC 90028	29
<i>C. krusei</i> ATCC 6258	23
<i>C. parapsilosis</i> ATCC 22019	26
<i>C. tropicalis</i> ATCC 750	31

LOTTO/BATCH : Idoneo / Approved

Non Idoneo / Not approved

DATA / DATE 21.02.2020

Responsabile Controllo Qualità
Quality Control Manager
(D. Vitagliano)

Dario Vitagliano