

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Tylosin TY 30 µg	101819120	2022.10.17

Ref.

9082 – 9082/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	18-26	22

Batch Release

Approved

Date

23.10.2019

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.