

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Tiamulin T 30 µg	060721069	2024.06.06

Ref.

9094 – 9094/1 – 9094/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	25-32	30

Batch Release

Approved

Date

11.06.2021

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.