

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Expiration date |
|------------------|-----------|-----------------|
| Tiamulin T 30 µg | 092421117 | 2024.09.23 |

Ref.

9094 – 9094/1 – 9094/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains | Medium | Inoculum | Incubation | Expected Results Zone range (mm) | Results Zone (mm) |
|---|---------------------------|---------------|------------------------------|-------------------------------------|----------------------|
| <i>Staphylococcus aureus</i> ATCC® 25923 | Mueller Hinton II Agar | 0.5 McFarland | 35 ± 1°C, ambient 16-18 h | 25-32 | 30 |

Batch Release

Approved

Date

30.09.2021

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.