

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Ticarcillin - clavulanic acid TTC 85 µg	062618068	2021.06.25

Ref.

9096 – 9096/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	24-30	28
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	29-37	31
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	21-25	25
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	20-28	27

Batch Release

Approved

Date

03.07.2018

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.