

|                             |                                    |   |
|-----------------------------|------------------------------------|---|
| <b>Liofilchem</b><br>S.r.l. | CERTIFICATO CONTROLLO QUALITÀ      | N°9097 - 9097/1                           |
|                             | <i>QUALITY CONTROL CERTIFICATE</i> | Revisione 7 del 09.10.2013<br>Pag. 1 di 1 |

**CARATTERISTICHE DEL PRODOTTO / PRODUCT CHARACTERISTICS**

|   |                           |
|---|---------------------------|
| Prodotto / <i>Product</i>   | <b>CLOTRIMAZOLE</b>       |
| Sigla / <i>Initial</i>  | <b>CLO 50 ug</b>          |
| Lotto / <i>Batch</i>  | <b>051617088</b>          |
| Data produzione / <i>Production date</i>                            | <b>16.05.2017</b>         |
| Data scadenza / <i>Expiry date</i>                                  | <b>2020.05.15</b>         |
| Contenuto antibiotico del disco / <i>Antimicrobial Disk content</i> | <b>Clotrimazole 50 ug</b> |

**CONTROLLO QUALITÀ / QUALITY CONTROL**

|  |   |    |
|--|---|----|
| Terreno di coltura / <i>Culture medium</i>               | <b>Mueller Hinton+Glucose+Methylen Blue Agar</b>  |    |
| Tempo di incubazione / <i>Incubation time</i>            | <b>20-24 h</b><br><i>in caso di crescita insufficiente prolungare l'incubazione fino a 48 ore / read at 48 hours only when insufficient growth is observed after 24 hours of incubation</i> |    |
| Temperatura / <i>Temperature</i>                         | <b>36 ± 1°C</b>   |    |
| Modalità di incubazione / <i>Procedure of incubation</i> | <b>O<sub>2</sub></b>  |    |
| Microrganismo/ <i>Microorganism</i>                      | <b>RESULTS</b><br>mm  |    |
| <i>C. albicans</i>                                       | ATCC 90028  | 28 |
| <i>C. krusei</i>   | ATCC 6258   | 35 |
| <i>C. parapsilosis</i>                                   | ATCC 22019  | 37 |
| <i>C. tropicalis</i>                                     | ATCC 750  | 31 |

LOTTO / BATCH :  Idoneo / Approved

Non Idoneo / Not approved

DATA / DATE : 24.05.2017

Responsabile Controllo Qualità  
*Quality Control Manager*  
(D. Vitagliano)

*Dario Vitagliano*