

<b>Liofilchem</b> S.r.l.	CERTIFICATO CONTROLLO QUALITÀ	N°9097 - 9097/1
	<i>QUALITY CONTROL CERTIFICATE</i>	Revisione 7 del 09.10.2013 Pag. 1 di 1

**CARATTERISTICHE DEL PRODOTTO / PRODUCT CHARACTERISTICS**

Prodotto / <i>Product</i>	<b>CLOTRIMAZOLE</b>
Sigla / <i>Initial</i>	<b>CLO 50 ug</b>
Lotto / <i>Batch</i>	<b>071717041</b>
Data produzione / <i>Production date</i>	<b>17.07.2017</b>
Data scadenza / <i>Expiry date</i>	<b>2020.07.16</b>
Contenuto antibiotico del disco / <i>Antimicrobial Disk content</i>	<b>Clotrimazole 50 ug</b>

**CONTROLLO QUALITÀ / QUALITY CONTROL**

Terreno di coltura / <i>Culture medium</i>	<b>Mueller Hinton+Glucose+Methylen Blue Agar</b>	
Tempo di incubazione / <i>Incubation time</i>	<b>20-24 h</b> <i>in caso di crescita insufficiente prolungare l'incubazione fino a 48 ore / read at 48 hours only when insufficient growth is observed after 24 hours of incubation</i>	
Temperatura / <i>Temperature</i>	<b>36 ± 1°C</b>	
Modalità di incubazione / <i>Procedure of incubation</i>	<b>O<sub>2</sub></b>	
Microrganismo/ <i>Microorganism</i>	<b>RESULTS</b> mm	
<i>C. albicans</i>	ATCC 90028	32
<i>C. krusei</i>	ATCC 6258	34
<i>C. parapsilosis</i>	ATCC 22019	30
<i>C. tropicalis</i>	ATCC 750	30

LOTTO / BATCH :  Idoneo / Approved

Non Idoneo / Not approved

DATA / DATE : 21.07.2017

Responsabile Controllo Qualità  
*Quality Control Manager*  
(D. Vitagliano)

*Dario Vitagliano*