

# Liofilchem®

## Certificate of Analysis

Product	Batch	Expiration date
Fosfomicin FOS 200 µg	052919098	2022.05.28

**Ref.**

9109 – 9109/1

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	22-30	29
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	26-34*	29
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	25-33	33

\*Established and validated by EUCAST

**Batch Release**

Approved

**Date**

03.06.2019

**Signature**Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.