

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Fosfomycin FOS 200 µg	052421040	2024.05.23

Ref.

9109 – 9109/1 – 9109/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	22-30	28
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	26-34*	28
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	25-33	27

*Established and validated by EUCAST

Batch Release

Approved

Date

01.06.2021

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.