

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Trimethoprim TM 5 µg	082318041	2021.08.22

Ref.

9110 – 9110/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	21-28	21
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	19-26	22
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	22-28*	22
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	24-32*	26


*Established and validated by EUCAST

Batch Release

Approved

Date

29.08.2018

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.