

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Gentamicin CN 120 µg	082319015	2022.08.22

Ref.

9124 – 9124/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	16-23	16
<i>Enterococcus faecalis</i> ATCC® 51299	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	6	6

Batch Release

Approved

Date

05.09.2019

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.