

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Mecillinam MEC 10 µg	111821081	2024.11.17

### Ref.

9156 – 9156/1 – 9156/2

### Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-20 h	24-30*	25

\*Established and validated by EUCAST

### Batch Release

Approved

Date

24.11.2021

Signature

Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.