

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Mupirocin MUP 200 µg	012020083	2023.01.19

Ref.

9157 – 9157/1

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	29–38	32
<i>Staphylococcus aureus</i> ATCC®-BAA 1708	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	6	6
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	31–37*	32

*Established and validated by EUCAST

Batch Release

Approved

Date

27.01.2020

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.