

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Mupirocin MUP 200 µg	120420122	2023.12.04

### Ref.

9157 – 9157/1

### Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	29–38	34
<i>Staphylococcus aureus</i> ATCC®-BAA 1708	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	6	6
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	31–37*	37

\*Established and validated by EUCAST

### Batch Release

Approved

Date

10.12.2020

Signature

Quality Control  
(D. Vitagliano)



The results reported were obtained at the time of release.