

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Mupirocin MUP 200 µg	011522001	2025.01.14

Ref.

9157 – 9157/1 – 9157/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	29–38	29
<i>Staphylococcus aureus</i> ATCC®-BAA 1708	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	6	6
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	31–37*	32

*Established and validated by EUCAST

Batch Release

Approved

Date

20.01.2022

Signature

Quality Control

(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.