

Liofilchem®

Certificate of Analysis

| Product | Batch | Expiration date |
|----------------------|-----------|-----------------|
| Caspofungin CAS 5 µg | 011422012 | 2025.01.13 |

Ref.

9165 – 9165/1 – 9165/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

| Control strains | Medium | Inoculum | Incubation | Results Halo (mm) |
|--|--|---------------|---------------------------------|----------------------|
| <i>Candida albicans</i> ATCC® 90028 | Mueller Hinton+ Glucose+ Methylene Blue Agar | McFarland 0.5 | Aerobic , 24-48 h , 35 ± 2°C | 18-27 22 |
| <i>Candida parapsilosis</i> ATCC® 22019 | Mueller Hinton+ Glucose+ Methylene Blue Agar | McFarland 0.5 | Aerobic , 24-48 h , 35 ± 2°C | 14-23 15 |
| <i>Candida krusei</i> ATCC® 6258 | Mueller Hinton+ Glucose+ Methylene Blue Agar | McFarland 0.5 | Aerobic , 24-48 h , 35 ± 2°C | 19-26 22 |
| <i>Candida tropicalis</i> ATCC® 750 | Mueller Hinton+ Glucose+ Methylene Blue Agar | McFarland 0.5 | Aerobic , 24-48 h , 35 ± 2°C | 20-27 26 |

Batch Release

Approved

Date

20.01.2022

SignatureQuality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.